
Midmark M11 - v1.0.0
Total Cycles: 3250

Name: RITTER 1
Sterilizer ID: V2178198
05/17/2023 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.0 F 0.0
2:00 190.4 F 2.5
4:00 198.6 F 7.5
6:00 220.8 F 12.8
8:00 230.0 F 8.2
10:00 241.0 F 12.0
12:00 241.0 F 11.5
14:00 239.3 F 10.1
16:00 253.2 F 17.0
18:00 264.9 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 272.4 F 28.8
3:00 271.5 F 28.2
4:00 271.9 F 28.9

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 233.5 F 0.0
10:00 228.4 F 0.0
15:00 227.5 F 0.0
20:00 226.3 F 0.0
25:00 229.7 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

05/17/2023 12:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____