

-----  
Midmark M11 - v1.0.0  
Total Cycles: 903

Name: RITTER 1  
Sterilizer ID: V2178198  
06/16/2021 06:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.8 F 0.0  
2:00 80.0 F 0.0  
4:00 83.5 F 0.0  
6:00 100.4 F 0.0  
8:00 136.6 F 0.1  
10:00 174.2 F 0.2  
12:00 195.5 F 1.6  
14:00 216.1 F 4.6  
16:00 223.2 F 7.2  
18:00 232.2 F 10.4  
20:00 235.5 F 9.6  
22:00 238.3 F 10.6  
24:00 235.6 F 8.2  
26:00 244.8 F 12.2  
28:00 245.2 F 12.8  
30:00 257.4 F 19.2  
32:00 268.2 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.3 F 28.5  
2:00 272.1 F 28.6  
3:00 272.0 F 28.3  
4:00 272.3 F 28.9  
  
Min 271.5 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.6 F 0.0  
5:00 178.2 F 0.0  
10:00 214.7 F 0.0  
15:00 222.5 F 0.0  
20:00 220.3 F 0.0  
25:00 216.0 F 0.0  
30:00 210.7 F 0.0  
  
Drying Complete

FILLING: 0:55  
HEATING: 33:16  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:11:29  
CYCLE COMPLETE

06/16/2021 08:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_