

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3394

Name: RITTER 1  
Sterilizer ID: V2178198  
06/15/2023 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.9 F 0.0  
2:00 189.9 F 0.5  
4:00 198.9 F 4.7  
6:00 219.1 F 9.5  
8:00 233.4 F 9.1  
10:00 234.5 F 10.0  
12:00 238.7 F 8.6  
14:00 243.4 F 12.9  
16:00 242.1 F 11.5  
18:00 255.2 F 18.2  
20:00 266.5 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 29.0  
2:00 271.5 F 28.5  
3:00 272.6 F 29.1  
4:00 271.7 F 28.4  
  
Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
5:00 212.0 F 0.0  
10:00 233.1 F 0.0  
15:00 230.4 F 0.0  
20:00 224.3 F 0.0  
25:00 218.5 F 0.0  
30:00 217.1 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

06/15/2023 10:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_