
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002336
14:08 29/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:59
STERILIZING 2:53
136.0 C 220kPa 2:53
136.3 C 221kPa 3:23
136.4 C 226kPa 3:53
136.4 C 224kPa 4:23
136.4 C 225kPa 4:53
136.4 C 225kPa 5:23
136.3 C 226kPa 5:53
136.4 C 226kPa 6:23
136.4 C 226kPa 6:23
VENTING 6:23
AIR DRYING 7:14
CYCLE COMPLETE 37:14

Digital Signature #
28DDDAE29535BBB
