
Midmark M11 - v1.0.5
Total Cycles: 3185

Name: RITTER2
Sterilizer ID: V2313228
12/05/2023 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.5 F 0.0
2:00 123.1 F 0.6
4:00 139.1 F 3.1
6:00 163.3 F 6.5
8:00 188.2 F 10.4
10:00 209.3 F 6.2
12:00 214.6 F 8.0
14:00 229.0 F 12.7
16:00 235.4 F 9.6
18:00 237.4 F 9.8
20:00 245.8 F 13.6
22:00 256.8 F 19.6
24:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.5
1:00 272.6 F 30.1
2:00 271.7 F 29.4
3:00 272.5 F 30.1
4:00 271.7 F 29.5

Min 271.3 F 29.3
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 206.5 F 0.0
10:00 192.5 F 0.0
15:00 189.5 F 0.0
20:00 186.9 F 0.0
25:00 184.4 F 0.0
30:00 182.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:29
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:04
CYCLE COMPLETE

12/05/2023 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____