

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3230

Name: RITTER 1  
Sterilizer ID: V2178198  
05/15/2023 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.4 F 0.0  
2:00 173.1 F 0.0  
4:00 189.8 F 2.4  
6:00 206.4 F 7.5  
8:00 221.0 F 12.6  
10:00 228.2 F 7.8  
12:00 240.1 F 13.1  
14:00 239.4 F 10.8  
16:00 237.4 F 9.1  
18:00 251.7 F 16.0  
20:00 263.4 F 22.8  
22:00 271.9 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.2  
1:00 272.2 F 28.8  
2:00 272.2 F 28.5  
3:00 271.5 F 28.3  
4:00 272.6 F 28.9

Min 271.4 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 216.8 F 0.0  
10:00 235.9 F 0.0  
15:00 229.6 F 0.0  
20:00 224.2 F 0.0  
25:00 219.5 F 0.0  
30:00 217.0 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 22:01  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:08  
CYCLE COMPLETE

05/15/2023 10:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_