
Midmark M11 - v1.0.0
Total Cycles: 1455

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2021 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.2 F 0.0
2:00 197.0 F 2.8
4:00 201.2 F 8.0
6:00 218.0 F 11.2
8:00 236.2 F 8.6
10:00 235.5 F 10.4
12:00 244.4 F 12.0
14:00 241.6 F 10.5
16:00 255.1 F 17.4
18:00 266.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 273.0 F 28.8
2:00 271.6 F 28.0
3:00 272.5 F 28.9
4:00 272.6 F 28.6

Min 271.1 F 28.0
Max 273.1 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 241.7 F 0.0
10:00 230.9 F 0.0
15:00 228.0 F 0.0
20:00 227.7 F 0.0
25:00 229.7 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

12/14/2021 12:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____