

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002419  
11:09 13/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:28  
STERILIZING 3:33  
136.0 C 222kPa 3:33  
136.4 C 226kPa 4:03  
136.2 C 223kPa 4:33  
136.1 C 223kPa 5:03  
136.1 C 223kPa 5:33  
136.3 C 226kPa 6:03  
136.3 C 225kPa 6:33  
136.2 C 224kPa 7:03  
136.2 C 224kPa 7:03  
VENTING 7:03  
AIR DRYING 7:55  
CYCLE COMPLETE 37:55  
-----

Digital Signature #  
F236AB1726FBE38B  
-----