
Midmark M11 - v1.0.0
Total Cycles: 3603

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.2 F 0.0
2:00 206.0 F 4.2
4:00 213.5 F 9.4
6:00 225.0 F 6.6
8:00 234.3 F 10.5
10:00 236.2 F 7.3
12:00 242.9 F 11.6
14:00 245.8 F 13.1
16:00 258.5 F 20.0
18:00 270.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.8
2:00 271.8 F 28.8
3:00 272.2 F 28.7
4:00 271.8 F 28.7

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 214.9 F 0.0
10:00 239.6 F 0.0
15:00 232.7 F 0.0
20:00 225.5 F 0.0
25:00 224.4 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 2:53
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

08/01/2023 03:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____