

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5061

Name: RITTER 1  
Sterilizer ID: V2178198  
11/05/2024 06:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.0 F 0.0  
2:00 185.9 F 4.8  
4:00 214.4 F 11.0  
6:00 224.3 F 7.1  
8:00 237.6 F 13.3  
10:00 239.9 F 11.2  
12:00 238.7 F 10.0  
14:00 253.8 F 17.4  
16:00 266.4 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.5 F 29.1  
2:00 271.5 F 28.5  
3:00 272.4 F 28.8  
4:00 271.7 F 28.7  
  
Min 271.4 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 221.9 F 0.0  
10:00 229.1 F 0.0  
15:00 233.6 F 0.0  
20:00 227.5 F 0.0  
25:00 226.1 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 17:21  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:56:28  
CYCLE COMPLETE

11/05/2024 06:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_