
Midmark M11 - v1.0.5
Total Cycles: 1911

Name: RITTER2
Sterilizer ID: V2313228
01/10/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.7 F 0.0
2:00 182.1 F 2.7
4:00 191.7 F 7.2
6:00 212.4 F 12.6
8:00 222.0 F 7.6
10:00 234.7 F 13.1
12:00 237.6 F 10.5
14:00 235.6 F 8.8
16:00 249.9 F 15.6
18:00 261.3 F 22.4
20:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 271.9 F 29.4
2:00 271.2 F 29.2
3:00 271.9 F 29.5
4:00 271.1 F 28.9

Min 271.1 F 28.9
Max 272.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 220.2 F 0.0
10:00 243.1 F 0.0
15:00 240.9 F 0.0
20:00 228.3 F 0.0
25:00 222.4 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

01/10/2023 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____