
Midmark M11 - v1.0.5
Total Cycles: 1743

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.0 F 0.0
2:00 165.1 F 0.2
4:00 204.4 F 1.1
6:00 218.5 F 5.7
8:00 231.9 F 10.9
10:00 231.1 F 7.4
12:00 244.0 F 13.5
14:00 243.2 F 12.3
16:00 242.7 F 11.9
18:00 255.9 F 19.0
20:00 267.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.6
2:00 271.8 F 28.8
3:00 272.5 F 29.8
4:00 272.6 F 29.3

Min 271.4 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 211.9 F 0.0
10:00 231.2 F 0.0
15:00 238.6 F 0.0
20:00 225.4 F 0.0
25:00 231.0 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

11/17/2022 02:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____