

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3135

Name: RITTER 1  
Sterilizer ID: V2178198  
04/20/2023 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
2:00 198.0 F 0.0  
4:00 214.5 F 3.5  
6:00 226.1 F 8.7  
8:00 237.1 F 10.9  
10:00 241.1 F 11.6  
12:00 239.9 F 10.2  
14:00 238.2 F 9.3  
16:00 253.7 F 16.9  
18:00 266.2 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.3 F 28.3  
2:00 271.9 F 28.3  
3:00 272.1 F 28.2  
4:00 271.8 F 28.2  
  
Min 271.4 F 27.7  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 221.1 F 0.0  
10:00 241.0 F 0.0  
15:00 238.6 F 0.0  
20:00 224.3 F 0.0  
25:00 220.7 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

04/20/2023 10:34 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_