
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002029
14:44 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:56
STERILIZING 2:56
136.4 C 224kPa 2:56
136.3 C 225kPa 3:27
136.2 C 225kPa 3:57
136.4 C 227kPa 4:27
136.2 C 224kPa 4:57
136.3 C 225kPa 5:27
136.2 C 224kPa 5:57
136.2 C 224kPa 6:27
136.2 C 224kPa 6:27
VENTING 6:27
AIR DRYING 7:17
CYCLE COMPLETE 37:17

Digital Signature #
6F2845E81CEB20B2
