
Midmark M11 - v1.0.5
Total Cycles: 3799

Name: RITTER2
Sterilizer ID: V2313228
07/19/2024 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 72.1 F 0.0
4:00 89.8 F 0.0
6:00 112.7 F 0.1
8:00 148.1 F 1.3
10:00 174.5 F 4.2
12:00 189.0 F 7.6
14:00 203.2 F 11.2
16:00 213.0 F 4.4
18:00 224.9 F 7.9
20:00 235.4 F 12.1
22:00 234.5 F 8.8
24:00 241.5 F 11.1
26:00 245.7 F 13.2
28:00 257.4 F 19.5
30:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.0
2:00 271.8 F 28.9
3:00 272.2 F 28.8
4:00 272.0 F 29.1

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.5 F 0.0
5:00 179.0 F 0.0
10:00 211.8 F 0.0
15:00 203.6 F 0.0
20:00 200.3 F 0.0
25:00 199.8 F 0.0
30:00 199.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

07/19/2024 09:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____