
Midmark M11 - v1.0.0
Total Cycles: 3946

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2023 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.2 F 0.0
2:00 132.8 F 1.7
4:00 144.0 F 4.7
6:00 177.0 F 8.8
8:00 204.8 F 9.5
10:00 218.7 F 7.7
12:00 229.0 F 12.8
14:00 233.4 F 9.1
16:00 239.4 F 10.0
18:00 245.1 F 13.0
20:00 257.4 F 19.5
22:00 268.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 29.1
2:00 271.2 F 28.4
3:00 272.8 F 29.2
4:00 271.8 F 28.6

Min 271.2 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 222.8 F 0.0
10:00 233.6 F 0.0
15:00 230.0 F 0.0
20:00 228.3 F 0.0
25:00 228.8 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

10/25/2023 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____