
Midmark M11 - v1.0.5
Total Cycles: 3476

Name: RITTER2
Sterilizer ID: V2313228
03/26/2024 09:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 72.5 F 0.0
4:00 94.9 F 0.1
6:00 112.4 F 0.7
8:00 137.0 F 2.8
10:00 163.8 F 6.0
12:00 182.3 F 9.7
14:00 202.0 F 11.0
16:00 216.0 F 5.8
18:00 225.3 F 9.6
20:00 234.8 F 10.9
22:00 237.9 F 10.8
24:00 235.2 F 8.3
26:00 248.2 F 14.7
28:00 259.1 F 20.8
30:00 269.3 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.4
2:00 271.4 F 28.8
3:00 272.6 F 29.5
4:00 271.6 F 28.7

Min 271.3 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 199.9 F 0.0
10:00 193.9 F 0.0
15:00 186.2 F 0.0
20:00 181.7 F 0.0
25:00 181.2 F 0.0
30:00 179.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:13
CYCLE COMPLETE

03/26/2024 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____