

-----  
Midmark M11 - v1.0.0  
Total Cycles: 930

Name: RITTER 1  
Sterilizer ID: V2178198  
06/23/2021 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
2:00 191.9 F 3.0  
4:00 197.9 F 6.8  
6:00 221.2 F 11.6  
8:00 224.0 F 6.0  
10:00 236.6 F 11.0  
12:00 233.4 F 7.8  
14:00 243.6 F 12.1  
16:00 245.1 F 12.7  
18:00 256.9 F 19.0  
20:00 267.8 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.6 F 28.7  
2:00 271.3 F 28.3  
3:00 272.5 F 28.8  
4:00 271.3 F 28.2  
  
Min 271.2 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 215.6 F 0.0  
10:00 212.4 F 0.0  
15:00 211.6 F 0.0  
20:00 210.3 F 0.0  
25:00 232.5 F 0.0  
30:00 243.7 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 21:22  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

06/23/2021 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_