
Midmark M11 - v1.0.5
Total Cycles: 4501

Name: RITTER2
Sterilizer ID: V2313228
04/10/2025 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 87.0 F 0.1
4:00 91.3 F 1.2
6:00 119.3 F 3.7
8:00 149.7 F 6.7
10:00 168.5 F 10.0
12:00 199.4 F 13.6
14:00 214.6 F 6.4
16:00 226.3 F 10.4
18:00 233.0 F 8.1
20:00 237.9 F 10.3
22:00 236.3 F 8.7
24:00 246.4 F 13.6
26:00 257.3 F 19.6
28:00 267.5 F 26.0
30:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 271.5 F 29.3
2:00 271.5 F 29.3
3:00 271.4 F 29.2
4:00 271.5 F 29.2

Min 271.3 F 29.2
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 233.0 F 0.0
10:00 198.3 F 0.0
15:00 209.8 F 0.0
20:00 188.0 F 0.0
25:00 175.9 F 0.0
30:00 174.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

04/10/2025 10:44 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____