
Midmark M11 - v1.0.5
Total Cycles: 3772

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 05:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.7 F 0.0
2:00 105.5 F 0.1
4:00 135.0 F 1.2
6:00 166.5 F 4.2
8:00 181.8 F 7.5
10:00 199.8 F 11.3
12:00 213.6 F 4.4
14:00 223.7 F 8.0
16:00 234.2 F 12.4
18:00 235.1 F 9.4
20:00 239.6 F 10.3
22:00 246.3 F 13.8
24:00 257.6 F 20.0
26:00 268.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 273.1 F 29.8
2:00 271.9 F 29.0
3:00 272.3 F 29.7
4:00 272.7 F 29.6

Min 271.5 F 28.9
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 212.8 F 0.0
10:00 225.5 F 0.0
15:00 216.0 F 0.0
20:00 213.2 F 0.0
25:00 211.4 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:06:09
CYCLE COMPLETE

07/11/2024 06:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____