
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000574
16:43 09/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 5:08
STERILIZING 6:07
136.4 C 231kPa 6:07
136.4 C 233kPa 6:37
136.1 C 229kPa 7:07
136.4 C 232kPa 7:37
136.3 C 231kPa 8:07
136.4 C 232kPa 8:37
136.2 C 230kPa 9:07
136.4 C 232kPa 9:37
136.2 C 231kPa 10:07
136.4 C 233kPa 10:37
136.4 C 232kPa 11:07
136.4 C 233kPa 11:37
136.2 C 232kPa 12:07
136.1 C 231kPa 12:37
136.4 C 233kPa 13:07
136.4 C 233kPa 13:37
136.3 C 231kPa 14:07
136.4 C 233kPa 14:37
136.3 C 231kPa 15:07
136.3 C 232kPa 15:37
136.2 C 232kPa 16:07
136.2 C 232kPa 16:07
VENTING 16:07
AIR DRYING 16:59
CYCLE COMPLETE 31:59

Digital Signature #
EA91B8065A85D399
