

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1143

Name: RITTER2  
Sterilizer ID: V2313228  
05/31/2022 09:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
2:00 188.3 F 3.1  
4:00 189.4 F 8.1  
6:00 208.1 F 10.0  
8:00 226.4 F 9.0  
10:00 235.2 F 10.5  
12:00 240.8 F 12.3  
14:00 240.3 F 11.3  
16:00 253.8 F 18.1  
18:00 265.7 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.2  
1:00 272.3 F 29.9  
2:00 272.5 F 29.6  
3:00 271.5 F 29.3  
4:00 272.7 F 30.1  
  
Min 271.4 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 200.3 F 0.0  
10:00 206.9 F 0.0  
15:00 210.8 F 0.0  
20:00 209.2 F 0.0  
25:00 207.8 F 0.0  
30:00 201.6 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

05/31/2022 10:05 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02