
Midmark M11 - v1.0.0
Total Cycles: 2776

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 158.5 F 0.0
4:00 179.3 F 2.0
6:00 203.1 F 6.6
8:00 220.4 F 11.4
10:00 225.1 F 6.6
12:00 238.9 F 12.1
14:00 237.6 F 9.7
16:00 236.7 F 8.7
18:00 250.2 F 15.3
20:00 262.4 F 22.3
22:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 29.1
2:00 271.8 F 28.4
3:00 271.4 F 28.4
4:00 272.5 F 29.1

Min 271.2 F 28.1
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 216.7 F 0.0
10:00 227.7 F 0.0
15:00 231.0 F 0.0
20:00 233.9 F 0.0
25:00 230.5 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:22
CYCLE COMPLETE

01/12/2023 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____