
Midmark M11 - v1.0.5
Total Cycles: 4182

Name: RITTER2
Sterilizer ID: V2313228
12/04/2024 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.7 F 0.0
2:00 146.4 F 0.2
4:00 156.8 F 1.0
6:00 183.8 F 4.8
8:00 200.6 F 9.0
10:00 227.3 F 13.4
12:00 223.9 F 8.6
14:00 236.5 F 12.7
16:00 238.9 F 11.6
18:00 239.2 F 10.6
20:00 251.6 F 17.1
22:00 262.7 F 23.8
24:00 272.6 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 273.1 F 30.4
2:00 272.5 F 29.8
3:00 271.7 F 29.5
4:00 272.8 F 30.4

Min 271.7 F 29.4
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 236.2 F 0.0
10:00 220.9 F 0.0
15:00 214.0 F 0.0
20:00 195.1 F 0.0
25:00 199.0 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:04
CYCLE COMPLETE

12/04/2024 02:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____