

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002322  
12:17 21/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:04  
STERILIZING 3:10  
136.4 C 224kPa 3:10  
136.3 C 224kPa 3:40  
136.2 C 224kPa 4:10  
136.1 C 222kPa 4:40  
136.3 C 223kPa 5:10  
136.5 C 226kPa 5:40  
136.1 C 224kPa 6:10  
136.2 C 224kPa 6:40  
136.2 C 224kPa 6:40  
VENTING 6:40  
AIR DRYING 7:32  
CYCLE COMPLETE 37:32

-----  
Digital Signature #  
1C50431776DB94D3  
-----