
Midmark M11 - v1.0.0
Total Cycles: 1483

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2021 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.4 F 0.0
2:00 106.8 F 0.0
4:00 128.8 F 0.1
6:00 163.0 F 1.3
8:00 193.3 F 4.3
10:00 208.2 F 8.2
12:00 221.8 F 12.4
14:00 226.1 F 6.8
16:00 238.0 F 11.8
18:00 235.6 F 9.0
20:00 239.4 F 9.6
22:00 247.7 F 14.0
24:00 260.3 F 21.0
26:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.3
1:00 272.3 F 28.9
2:00 272.4 F 28.6
3:00 271.9 F 28.6
4:00 272.8 F 28.9

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.8 F 0.0
5:00 213.3 F 0.0
10:00 227.5 F 0.0
15:00 227.7 F 0.0
20:00 227.3 F 0.0
25:00 229.7 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 26:28
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

12/21/2021 04:26 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____