
Midmark M11 - v1.0.0
Total Cycles: 1677

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.4 F 0.0
2:00 167.9 F 1.7
4:00 174.9 F 5.2
6:00 196.2 F 10.0
8:00 215.3 F 5.1
10:00 229.1 F 9.3
12:00 236.5 F 9.5
14:00 242.1 F 12.1
16:00 240.8 F 10.5
18:00 254.3 F 17.4
20:00 266.2 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.6 F 28.1
3:00 272.6 F 29.0
4:00 272.2 F 28.5

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 226.1 F 0.0
10:00 233.4 F 0.0
15:00 237.3 F 0.0
20:00 224.7 F 0.0
25:00 223.0 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

02/28/2022 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____