
Midmark M11 - v1.0.0
Total Cycles: 4134

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2023 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.4 F 0.0
2:00 120.9 F 1.0
4:00 127.4 F 3.0
6:00 154.2 F 5.7
8:00 184.4 F 9.4
10:00 198.2 F 13.5
12:00 211.5 F 5.2
14:00 221.3 F 9.0
16:00 232.3 F 13.2
18:00 234.6 F 9.3
20:00 239.9 F 10.8
22:00 244.3 F 12.6
24:00 255.4 F 18.4
26:00 265.7 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.7 F 29.0
2:00 272.3 F 29.0
3:00 271.3 F 28.5
4:00 271.6 F 29.0

Min 271.0 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 217.0 F 0.0
10:00 220.9 F 0.0
15:00 230.1 F 0.0
20:00 222.0 F 0.0
25:00 231.4 F 0.0
30:00 232.0 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 27:44
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:28
CYCLE COMPLETE

12/19/2023 11:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____