

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4836

Name: RITTER 1  
Sterilizer ID: V2178198  
08/21/2024 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 74.6 F 0.0  
4:00 86.3 F 0.0  
6:00 119.3 F 0.2  
8:00 142.2 F 0.6  
10:00 174.3 F 1.4  
12:00 195.7 F 4.1  
14:00 206.3 F 7.2  
16:00 218.3 F 10.8  
18:00 222.5 F 6.2  
20:00 231.8 F 9.0  
22:00 241.6 F 12.3  
24:00 240.4 F 10.9  
26:00 236.4 F 8.7  
28:00 250.6 F 15.4  
30:00 262.0 F 21.8  
32:00 272.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.7  
2:00 271.5 F 28.0  
3:00 272.7 F 28.7  
4:00 271.7 F 28.0  
  
Min 271.5 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.5 F 0.0  
5:00 200.5 F 0.0  
10:00 204.5 F 0.0  
15:00 209.8 F 0.0  
20:00 205.5 F 0.0  
25:00 205.5 F 0.0  
30:00 208.7 F 0.0  
  
Drying Complete

FILLING: 0:40  
HEATING: 32:28  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:10:28  
CYCLE COMPLETE

08/21/2024 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_