
Midmark M11 - v1.0.5
Total Cycles: 268

Name: RITTER2
Sterilizer ID: V2313228
08/09/2021 06:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 71.4 F 0.1
4:00 113.9 F 0.9
6:00 115.1 F 3.2
8:00 139.5 F 6.3
10:00 170.3 F 10.3
12:00 210.9 F 4.9
14:00 216.8 F 7.0
16:00 228.2 F 11.0
18:00 228.9 F 7.1
20:00 240.6 F 12.2
22:00 237.0 F 10.0
24:00 250.2 F 16.3
26:00 261.2 F 22.8
28:00 271.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.8
1:00 272.2 F 29.8
2:00 271.9 F 29.9
3:00 272.0 F 30.2
4:00 272.1 F 29.7

Min 271.3 F 29.7
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 221.2 F 0.0
10:00 187.3 F 0.0
15:00 212.6 F 0.0
20:00 232.5 F 0.0
25:00 186.6 F 0.0
30:00 180.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:24
STERILIZING: 4:00
VENTING: 2:49
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

08/09/2021 07:43 AM
ACCEPTED: X
REJECTED:
Operator: U01