

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002270  
10:54 23/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:46  
STERILIZING 3:44  
136.4 C 223kPa 3:44  
136.1 C 222kPa 4:14  
136.1 C 219kPa 4:44  
136.4 C 223kPa 5:14  
136.3 C 223kPa 5:44  
136.4 C 223kPa 6:14  
136.4 C 222kPa 6:44  
136.2 C 222kPa 7:14  
136.2 C 222kPa 7:14  
VENTING 7:15  
CYCLE ABORTED 7:46  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 7:46  
-----

Digital Signature #  
FB4B7CF55CCD0EE4  
-----