

-----  
Midmark M11 - v1.0.0  
Total Cycles: 648

Name: MIDMARK  
Sterilizer ID: V2178198  
02/02/2021 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 73.8 F 0.0  
4:00 99.4 F 0.0  
6:00 138.8 F 0.1  
8:00 168.3 F 0.3  
10:00 184.8 F 2.6  
12:00 202.9 F 5.9  
14:00 222.9 F 9.5  
16:00 231.5 F 13.5  
18:00 230.2 F 8.0  
20:00 240.4 F 12.7  
22:00 237.8 F 9.7  
24:00 237.7 F 8.5  
26:00 248.9 F 14.6  
28:00 261.1 F 21.3  
30:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.1 F 28.4  
2:00 272.7 F 28.9  
3:00 271.5 F 28.1  
4:00 272.3 F 28.5

Min 271.5 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 223.0 F 0.0  
10:00 228.1 F 0.0  
15:00 226.3 F 0.0  
20:00 227.8 F 0.0  
25:00 232.6 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:32  
STERILIZING: 4:00  
VENTING: 2:52  
DRYING: 30:00  
TOTAL CYCLE: 01:08:20  
CYCLE COMPLETE

02/02/2021 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_