
Midmark M11 - v1.0.5
Total Cycles: 3555

Name: RITTER2
Sterilizer ID: V2313228
04/24/2024 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.3 F 0.0
2:00 136.5 F 1.9
4:00 151.6 F 4.8
6:00 184.3 F 8.8
8:00 205.2 F 13.4
10:00 219.4 F 7.4
12:00 229.3 F 12.4
14:00 232.3 F 9.4
16:00 237.9 F 9.6
18:00 246.3 F 14.2
20:00 258.1 F 20.7
22:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.3 F 30.1
2:00 271.3 F 29.6
3:00 272.2 F 30.3
4:00 271.6 F 29.6

Min 271.2 F 29.4
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 238.0 F 0.0
10:00 220.8 F 0.0
15:00 212.2 F 0.0
20:00 220.1 F 0.0
25:00 177.2 F 0.0
30:00 172.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:06
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

04/24/2024 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____