
Midmark M11 - v1.0.5
Total Cycles: 3701

Name: RITTER2
Sterilizer ID: V2313228
06/14/2024 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.2 F 0.0
2:00 119.0 F 1.0
4:00 134.4 F 3.3
6:00 159.9 F 6.4
8:00 181.5 F 9.8
10:00 198.3 F 11.8
12:00 222.2 F 6.9
14:00 230.7 F 11.1
16:00 230.8 F 7.2
18:00 240.2 F 12.3
20:00 239.0 F 10.4
22:00 250.8 F 16.2
24:00 260.7 F 22.2
26:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.6
1:00 271.3 F 29.7
2:00 272.1 F 30.0
3:00 272.1 F 29.8
4:00 272.1 F 29.6

Min 271.1 F 29.5
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 238.9 F 0.0
10:00 186.6 F 0.0
15:00 184.9 F 0.0
20:00 181.4 F 0.0
25:00 178.5 F 0.0
30:00 177.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:11
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:49
CYCLE COMPLETE

06/14/2024 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____