

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4373

Name: RITTER 1  
Sterilizer ID: V2178198  
03/08/2024 03:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.1 F 0.0  
2:00 155.5 F 2.9  
4:00 170.4 F 6.3  
6:00 190.7 F 10.0  
8:00 222.9 F 6.3  
10:00 222.6 F 7.8  
12:00 234.6 F 12.7  
14:00 236.0 F 9.4  
16:00 237.7 F 9.4  
18:00 247.2 F 13.7  
20:00 258.5 F 19.9  
22:00 268.7 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.5 F 29.3  
2:00 271.7 F 28.7  
3:00 272.5 F 29.0  
4:00 272.2 F 29.3

Min 271.0 F 28.2  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 239.4 F 0.0  
10:00 228.3 F 0.0  
15:00 228.9 F 0.0  
20:00 231.0 F 0.0  
25:00 227.6 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 23:07  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:01:57  
CYCLE COMPLETE

03/08/2024 04:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_