
Midmark M11 - v1.0.0
Total Cycles: 3668

Name: RITTER 1
Sterilizer ID: V2178198
08/17/2023 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 173.2 F 2.2
4:00 173.6 F 5.9
6:00 193.3 F 10.5
8:00 216.2 F 4.3
10:00 228.4 F 9.4
12:00 234.4 F 9.4
14:00 241.1 F 11.9
16:00 239.4 F 10.2
18:00 252.6 F 16.8
20:00 263.8 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.0 F 28.6
2:00 272.6 F 29.1
3:00 271.8 F 28.7
4:00 272.0 F 28.6

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 222.7 F 0.0
10:00 222.9 F 0.0
15:00 224.8 F 0.0
20:00 224.4 F 0.0
25:00 225.0 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:00
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

08/17/2023 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____