
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000693
8:09 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:43
PRESSURIZING 1:53
STERILIZING 3:01
136.4 C 233kPa 3:01
136.3 C 232kPa 3:31
136.4 C 233kPa 4:01
136.4 C 233kPa 4:31
136.4 C 233kPa 5:01
136.4 C 233kPa 5:31
136.3 C 232kPa 6:01
136.4 C 232kPa 6:31
136.4 C 232kPa 6:31
VENTING 6:31
AIR DRYING 7:23
CYCLE COMPLETE 7:55

Digital Signature #
568207E3DDBDF22F
