
Midmark M11 - v1.0.5
Total Cycles: 2668

Name: RITTER2
Sterilizer ID: V2313228
07/19/2023 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 183.0 F 3.3
4:00 197.6 F 8.0
6:00 217.3 F 13.0
8:00 228.0 F 8.6
10:00 239.4 F 11.9
12:00 239.9 F 11.6
14:00 239.3 F 10.5
16:00 252.3 F 17.0
18:00 263.6 F 23.5
20:00 272.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 271.7 F 29.5
2:00 271.7 F 29.4
3:00 272.1 F 29.5
4:00 272.6 F 29.7

Min 271.5 F 29.4
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 209.7 F 0.0
10:00 228.8 F 0.0
15:00 224.0 F 0.0
20:00 221.8 F 0.0
25:00 225.3 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:20
CYCLE COMPLETE

07/19/2023 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____