
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006155
14:23 17/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:10
STERILIZING 3:14
136.0 C 220kPa 3:14
136.4 C 224kPa 3:44
136.4 C 222kPa 4:14
136.3 C 224kPa 4:44
136.3 C 224kPa 5:14
136.4 C 225kPa 5:44
136.4 C 225kPa 6:14
136.3 C 220kPa 6:44
136.3 C 220kPa 6:44
VENTING 6:44
AIR DRYING 7:35
CYCLE COMPLETE 10:58

Digital Signature #
781234EBFD8AFD76
