
Midmark M11 - v1.0.5
Total Cycles: 1919

Name: RITTER2
Sterilizer ID: V2313228
01/11/2023 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 198.6 F 4.0
4:00 199.2 F 9.2
6:00 218.4 F 7.0
8:00 230.6 F 10.7
10:00 233.0 F 8.5
12:00 239.6 F 10.2
14:00 248.6 F 14.8
16:00 261.3 F 22.0
18:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.2 F 29.3
2:00 271.3 F 28.9
3:00 272.1 F 29.4
4:00 271.2 F 28.7

Min 271.2 F 28.7
Max 272.3 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 227.6 F 0.0
10:00 238.6 F 0.0
15:00 240.9 F 0.0
20:00 229.2 F 0.0
25:00 226.3 F 0.0
30:00 225.8 F 0.0

Drying Complete

FILLING: 1:58
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

01/11/2023 12:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____