
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001824
8:44 11/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:56
STERILIZING 4:04
136.0 C 221kPa 4:04
136.1 C 224kPa 4:34
136.1 C 222kPa 5:04
136.4 C 226kPa 5:34
136.3 C 224kPa 6:04
136.1 C 222kPa 6:34
136.4 C 225kPa 7:04
136.3 C 224kPa 7:34
136.3 C 224kPa 7:34
VENTING 7:34
AIR DRYING 8:22
CYCLE COMPLETE 37:28

Digital Signature #
CF874FCCD038754A
