

-----  
Midmark M11 - v1.0.5  
Total Cycles: 301

Name: RITTER2  
Sterilizer ID: V2313228  
08/23/2021 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.1 F 0.0  
2:00 177.1 F 3.9  
4:00 202.7 F 9.6  
6:00 216.6 F 6.7  
8:00 230.5 F 11.3  
10:00 235.4 F 9.4  
12:00 237.7 F 9.4  
14:00 250.7 F 16.1  
16:00 263.3 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.0 F 29.1  
2:00 272.6 F 29.7  
3:00 271.5 F 28.9  
4:00 272.4 F 29.4  
  
Min 271.5 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 230.7 F 0.0  
10:00 224.6 F 0.0  
15:00 207.6 F 0.0  
20:00 196.5 F 0.0  
25:00 190.7 F 0.0  
30:00 192.0 F 0.0  
Drying Complete

FILLING: 2:12  
HEATING: 17:59  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:18  
CYCLE COMPLETE

08/23/2021 11:31 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02