
Midmark M11 - v1.0.0
Total Cycles: 2831

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 192.4 F 2.9
4:00 194.9 F 7.2
6:00 206.5 F 12.1
8:00 221.1 F 6.6
10:00 235.6 F 11.8
12:00 234.8 F 8.7
14:00 240.8 F 10.2
16:00 246.9 F 13.4
18:00 259.2 F 20.2
20:00 270.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.4 F 28.9
2:00 271.8 F 28.3
3:00 271.2 F 28.3
4:00 272.4 F 29.0

Min 271.1 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 218.3 F 0.0
10:00 240.2 F 0.0
15:00 235.8 F 0.0
20:00 219.9 F 0.0
25:00 219.1 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:28
CYCLE COMPLETE

01/31/2023 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____