

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002425  
8:54 17/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:41  
STERILIZING 3:43  
136.0 C 220kPa 3:43  
136.2 C 223kPa 4:13  
136.2 C 221kPa 4:43  
136.0 C 222kPa 5:13  
136.4 C 225kPa 5:43  
136.4 C 225kPa 6:13  
136.4 C 224kPa 6:43  
136.4 C 224kPa 7:13  
136.4 C 224kPa 7:13  
VENTING 7:13  
AIR DRYING 8:02  
CYCLE COMPLETE 38:02

-----  
Digital Signature #  
530479D279EAA210  
-----