

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3249

Name: RITTER2  
Sterilizer ID: V2313228  
12/22/2023 10:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 170.8 F 1.4  
4:00 176.2 F 4.9  
6:00 196.3 F 9.4  
8:00 226.6 F 8.1  
10:00 231.2 F 8.7  
12:00 234.1 F 11.4  
14:00 241.3 F 11.4  
16:00 239.4 F 10.1  
18:00 252.2 F 16.5  
20:00 263.1 F 23.1  
22:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.3  
2:00 272.0 F 29.5  
3:00 272.1 F 29.1  
4:00 272.3 F 29.6

Min 271.3 F 28.8  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 232.7 F 0.0  
10:00 192.1 F 0.0  
15:00 184.8 F 0.0  
20:00 192.3 F 0.0  
25:00 188.5 F 0.0  
30:00 181.4 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 23:03  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:51  
CYCLE COMPLETE

12/22/2023 11:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_