

-----  
Midmark M11 - v1.0.5  
Total Cycles: 50

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2021 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.6 F 0.0  
2:00 130.6 F 1.5  
4:00 145.5 F 4.6  
6:00 170.7 F 8.4  
8:00 193.0 F 13.1  
10:00 211.2 F 6.3  
12:00 224.5 F 11.1  
14:00 230.6 F 8.0  
16:00 241.0 F 12.6  
18:00 243.0 F 12.7  
20:00 255.0 F 18.9  
22:00 265.6 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.3 F 29.9  
2:00 271.4 F 29.4  
3:00 272.5 F 30.0  
4:00 271.7 F 29.3

Min 271.4 F 29.3  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.0 F 0.0  
5:00 213.6 F 0.0  
10:00 211.4 F 0.0  
15:00 201.2 F 0.1  
20:00 207.6 F 0.0  
25:00 207.4 F 0.0  
30:00 209.2 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 23:45  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:02:27  
CYCLE COMPLETE

05/27/2021 01:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_