

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4106

Name: RITTER2  
Sterilizer ID: V2313228  
11/11/2024 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.7 F 0.0  
2:00 133.7 F 0.0  
4:00 141.7 F 0.4  
6:00 180.0 F 4.1  
8:00 204.0 F 8.3  
10:00 213.7 F 12.9  
12:00 221.1 F 7.6  
14:00 233.9 F 12.4  
16:00 234.6 F 9.7  
18:00 237.4 F 9.6  
20:00 246.6 F 14.3  
22:00 258.3 F 20.7  
24:00 268.8 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.1 F 29.5  
2:00 272.3 F 29.9  
3:00 271.8 F 29.9  
4:00 271.7 F 29.3

Min 271.4 F 29.2  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 214.9 F 0.0  
10:00 205.6 F 0.0  
15:00 196.4 F 0.0  
20:00 195.3 F 0.0  
25:00 192.2 F 0.0  
30:00 190.3 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 25:06  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:03:25  
CYCLE COMPLETE

11/11/2024 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_