
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003425
13:02 11/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:33
STERILIZING 3:34
136.5 C 221kPa 3:34
136.4 C 226kPa 4:04
136.2 C 224kPa 4:34
136.3 C 224kPa 5:04
136.4 C 225kPa 5:34
136.3 C 224kPa 6:04
136.1 C 220kPa 6:34
136.4 C 225kPa 7:04
136.4 C 225kPa 7:04
VENTING 7:05
AIR DRYING 7:54
CYCLE COMPLETE 7:57

Digital Signature #
31FB568D763C3A33
