
Midmark M11 - v1.0.5
Total Cycles: 2233

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 193.7 F 4.1
4:00 208.5 F 9.9
6:00 222.9 F 6.9
8:00 236.2 F 11.8
10:00 237.6 F 10.2
12:00 236.5 F 9.3
14:00 251.6 F 16.6
16:00 263.7 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.4
2:00 271.2 F 29.0
3:00 272.5 F 29.5
4:00 271.6 F 28.8

Min 271.1 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 231.9 F 0.0
10:00 220.3 F 0.0
15:00 218.9 F 0.0
20:00 217.4 F 0.0
25:00 223.4 F 0.0
30:00 230.1 F 0.0

Drying Complete

FILLING: 2:35
HEATING: 17:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

04/11/2023 03:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____