

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3734

Name: RITTER 1  
Sterilizer ID: V2178198  
09/07/2023 12:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
2:00 187.3 F 2.4  
4:00 204.2 F 6.5  
6:00 209.7 F 11.3  
8:00 224.3 F 6.1  
10:00 236.6 F 11.1  
12:00 233.8 F 8.2  
14:00 243.4 F 11.8  
16:00 245.2 F 12.9  
18:00 256.9 F 19.0  
20:00 267.3 F 25.4  
22:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.3 F 28.7  
2:00 272.3 F 29.0  
3:00 272.5 F 29.3  
4:00 271.8 F 28.9

Min 271.7 F 28.6  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 237.5 F 0.0  
10:00 228.4 F 0.0  
15:00 227.8 F 0.0  
20:00 228.3 F 0.0  
25:00 228.9 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 22:10  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:01:34  
CYCLE COMPLETE

09/07/2023 01:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_